



**Radon Test Report**

27-Aug-2020

<b>Customer Information:</b> Joe Tech 123 Main Street Frederick MD 21703	<b>Test Site:</b> 999 Somewhere Pl. Frederick MD 21703
---	--

A Rad Elec Recon® CRM (NRPP Device Code #8304 / NRSB Device Code #31823) was used for radon screening measurements that were conducted at the above referenced test site by: ABC Radon Testing

**The results are as follows:**

Serial#	Instrument	Location	Start Date/Time	End Date/Time	Results (pCi/L)
101	Recon CRM	Basement	08-25-2020 10:55	08-27-2020 10:55	3.0

**Average Radon Concentration in:                      Basement                      3.0 pCi/L**

**Analyzed By:** Joe Tech (NRSB #1234567890)                      Cal. Date: 9/23/2019    Cal. Due: 09/23/2020  
**Deployed By:** Joe Tech (NRSB #1234567890)  
**Retrieved By:** Joe Tech (NRSB #1234567890)

**Protocol:** Closed Building Conditions Met  
**Tampering:** No Tampering Detected  
**Weather:** No Abnormal Weather Conditions  
**Mitigation:** No Mitigation System Installed  
**Comment:** Thanks for the business!

**Radon Health Risk Information**

Radon is the second leading cause of lung cancer after smoking. The U.S. Environmental Protection Agency (US EPA) and the Surgeon General strongly recommend that further action be taken when a home’s radon test results are 4.0 pCi/L or greater. The national average indoor radon level is about 1.3 pCi/L. The higher the home’s radon level, the greater the health risk to you and your family. Reducing your radon levels can be done easily, effectively and fairly inexpensively. Even homes with very high radon levels can be reduced below 4.0 pCi/L. Please refer to the EPA website at [www.epa.gov/radon](http://www.epa.gov/radon) for further information to assist you in evaluating your test results or deciding if further action is needed.

Signature:                     Sample Signature                     

Date:                     Aug 27, 2020



**Graphical Radon Report**

27-Aug-2020


<b>Customer Information:</b> Joe Tech 123 Main Street Frederick MD 21703	<b>Test Site:</b> 999 Somewhere Pl. Frederick MD 21703
---	--

A Rad Elec Recon® CRM (NRPP Device Code #8304 / NRSB Device Code #31823) was used for radon screening measurements that were conducted at the above referenced test site by: ABC Radon Testing

**The results are as follows:**

Serial#	Instrument	Location	Start Date/Time	End Date/Time	Results (pCi/L)
101	Recon CRM	Basement	08-25-2020 10:55	08-27-2020 10:55	3.0

**Average Radon Concentration in: Basement 3.0 pCi/L**





**Hourly Radon Report**

27-Aug-2020

<b>Customer Information:</b> Joe Tech 123 Main Street Frederick MD 21703	<b>Test Site:</b> 999 Somewhere Pl. Frederick MD 21703
---	--

A Rad Elec Recon® CRM (NRPP Device Code #8304 / NRSB Device Code #31823) was used for radon screening measurements that were conducted at the above referenced test site by: ABC Radon Testing

**The results are as follows:**

Serial#	Instrument	Location	Start Date/Time	End Date/Time	Results (pCi/L)
101	Recon CRM	Basement	08-25-2020 10:55	08-27-2020 10:55	3.0

**Average Radon Concentration in: Basement 3.0 pCi/L**

Record#	Date/Time	Radon (pCi/L)	Temperature (°F)	Pressure (inHg)	Humidity (%)	Tilts
1	Aug-25-2020 11:55	2.3	77	29.1	48	0
2	Aug-25-2020 12:55	2.4	77	29.1	48	0
3	Aug-25-2020 13:55	2.7	78	29.1	48	0
4	Aug-25-2020 14:55	3.0	77	29.0	47	0
5	Aug-25-2020 15:55	3.6	77	29.0	47	0
6	Aug-25-2020 16:55	2.3	76	29.0	46	0
7	Aug-25-2020 17:55	2.7	76	29.0	46	0
8	Aug-25-2020 18:55	2.7	74	29.0	46	0
9	Aug-25-2020 19:55	2.8	74	29.0	46	0
10	Aug-25-2020 20:55	2.6	75	29.1	47	0
11	Aug-25-2020 21:55	3.4	75	29.1	47	0
12	Aug-25-2020 22:55	2.9	76	29.1	47	0
13	Aug-25-2020 23:55	2.8	76	29.1	48	0
14	Aug-26-2020 00:55	3.0	75	29.1	47	0
15	Aug-26-2020 01:55	3.6	75	29.1	48	0
16	Aug-26-2020 02:55	2.8	75	29.1	48	0
17	Aug-26-2020 03:55	3.6	75	29.1	48	0
18	Aug-26-2020 04:55	3.4	75	29.1	49	0
19	Aug-26-2020 05:55	3.4	75	29.1	49	0
20	Aug-26-2020 06:55	4.1	74	29.2	49	0
21	Aug-26-2020 07:55	3.4	74	29.2	50	0
22	Aug-26-2020 08:55	3.3	74	29.2	50	0
23	Aug-26-2020 09:55	4.8	74	29.2	50	0
24	Aug-26-2020 10:55	4.4	75	29.2	50	0
25	Aug-26-2020 11:55	3.7	76	29.2	50	0
26	Aug-26-2020 12:55	3.4	76	29.2	49	0
27	Aug-26-2020 13:55	2.8	77	29.2	48	0
28	Aug-26-2020 14:55	3.7	77	29.2	48	0
29	Aug-26-2020 15:55	2.8	77	29.1	47	0
30	Aug-26-2020 16:55	2.7	77	29.1	46	0
31	Aug-26-2020 17:55	2.9	76	29.1	46	0
32	Aug-26-2020 18:55	2.9	74	29.1	45	0



**Hourly Radon Report**

27-Aug-2020

<b>Customer Information:</b> Joe Tech 123 Main Street Frederick MD 21703	<b>Test Site:</b> 999 Somewhere Pl. Frederick MD 21703
---	--

A Rad Elec Recon® CRM (NRPP Device Code #8304 / NRSB Device Code #31823) was used for radon screening measurements that were conducted at the above referenced test site by: ABC Radon Testing

**The results are as follows:**

Serial#	Instrument	Location	Start Date/Time	End Date/Time	Results (pCi/L)
101	Recon CRM	Basement	08-25-2020 10:55	08-27-2020 10:55	3.0

**Average Radon Concentration in:                      Basement                      3.0 pCi/L**

Record#	Date/Time	Radon (pCi/L)	Temperature (°F)	Pressure (inHg)	Humidity (%)	Tilts
33	Aug-26-2020 19:55	2.7	74	29.1	45	0
34	Aug-26-2020 20:55	2.3	75	29.1	46	0
35	Aug-26-2020 21:55	2.6	75	29.1	47	0
36	Aug-26-2020 22:55	2.4	76	29.1	47	0
37	Aug-26-2020 23:55	2.6	76	29.1	48	0
38	Aug-27-2020 00:55	3.0	74	29.1	47	0
39	Aug-27-2020 01:55	2.3	75	29.1	47	0
40	Aug-27-2020 02:55	3.0	73	29.1	47	0
41	Aug-27-2020 03:55	2.7	74	29.0	48	0
42	Aug-27-2020 04:55	3.4	74	29.0	48	0
43	Aug-27-2020 05:55	2.6	74	29.1	49	0
44	Aug-27-2020 06:55	2.5	75	29.1	49	0
45	Aug-27-2020 07:55	3.4	75	29.1	49	0
46	Aug-27-2020 08:55	1.6	75	29.1	50	0
47	Aug-27-2020 09:55	2.1	75	29.1	50	0
48	Aug-27-2020 10:55	1.5	76	29.0	50	0